
# DRDC project results and Magnetic nanoparticle imaging


Yutong Zhao

Jan 28<sup>th</sup> 2019


# DRDC project results



# DRDC project results



## Summary of solutions



- We determined 5 ionic solutions and 1 molecular solution.
- Real part: all of these solutions decrease with the concentration (C).
- Imaginary part: NaCl, NH4NO<sub>3</sub> increase and Sugar drops with C.

#### Remaining problem:

• For gasoline, it has been stored 1 year. The high evaporation property may result the difference. Solution: get some fresh gasoline from gas station.

## Medical imaging methods


#### 1. Computed Tomography(CT)




Beer–Lambert law  $I = I_0 e^{-\alpha d}$ 

 $\alpha$  – attenuation factor


#### 2. Magnetic Resonance Imaging(MRI)




γ- gyromagnetic ratio




#### 3. Magnetic Particle Imaging(MPI)






Response at different magnetic field for Superparamagnetic nanoparticles




### MPI principle



MPI: high resolution

CT: X-ray
Sensitive to heavy elements:
Bones (Calcium)




MRI: Larmor procession Sensitive to light elements: Soft tissue (Hydrogen)


Carotid Phantom MPI image no Stenosis Carotid Phantom MPI image depicting stenosis with Stenosis 6 cm Z FOV a



## In vivo magnetic resonance imaging of hyperpolarized silicon particles


M. C. Cassidy<sup>1</sup>, H. R. Chan<sup>2</sup>, B. D. Ross<sup>2</sup>, P. K. Bhattacharya<sup>2,3</sup> and C. M. Marcus<sup>4,5</sup>\*







#### Microwave Cancer Imaging Exploiting Magnetic Nanoparticles as Contrast Agent

Gennaro Bellizzi, Ovidio M. Bucci\*, Fellow, IEEE, and Ilaria Catapano, Member, IEEE






#### What we can do for MPI?



Microwave imaging for nanoparticles

- Imaging without strong H field?
- Shallow in vivo imaging?



 $\mu$  material near field imaging

- Repeatability problem
- Re-design the sensor

